# PDL 828 "Sepia II"



# Computer Controlled Multichannel Diode Laser Driver

- Modular system, drives up to 8 laser or LED heads
- Pulsed and cw operation
- Computer controlled through USB
- Various operation modes: bursts, simultaneous, delayed, sequential, burst sequence
- Laser heads from 375 to 1990 nm, LED heads from 255 to 600 nm



two channel version



# Applications

- Multicolour time-resolved fluorescence spectroscopy (PIE)
- Diffuse optical tomography
- Multiwavelength laser ranging / LIDAR
- Molecular imaging
- Quantum optics, single photon generation

# **Computer Controlled Multichannel Diode Laser Driver**

The PDL 828 "Sepia II" is a completely computer controlled multichannel diode laser driver connected to the PC via USB. The PDL 828 "Sepia II" provides maximum flexibility for multiple wavelengths applications and drives any combination of up to 8 laser or LED heads in

parallel or in a user defined sequence. Laser heads with wavelengths between 375 and 1990 nm (LDH-P/D/FA Series) as well as pulsed LEDs from 255 to 600 nm (PLS Series) are available. The whole system can be configured and controlled through a dedicated Windows<sup>™</sup> control software. Last settings are saved inside the PDL 828 "Sepia II" to allow stand-alone operation making it a powerful device for measurement automation. A DLL is also available and allows to access all functions of the PDL 828 "Sepia II" from custom programs.

The system consists of a mainframe with power supply, an oscillator module and up to eight laser driver modules. The laser driver module is fully compatible to all LDH-P Series laser heads and can in addition drive the latest generation of laser heads, which also allow cw operation (LDH-D-C Series).

SLM 828

SYNC

OUT

FAST

GATE

SLOW

LASER

TRIG IN

STATUS

<u>0</u>	Sepia II	USB-Channe	l: O SerialNo.	: 1003603				
	Slot	000			Slo	t 100		
	SCM 828		SOM 828 4					
			Base Oscillator Divider		Mair	n Clock Pre Syn	: Sync Mask	
	O Po	wer	80.00 MHz (int.		3 80.000	0 MHz 0	Width	0 🕂
	💍 La:	ser Active	rising edge (ext falling edge (ext		Burst Out Burst	Length Sync En	able Invert 🗆	
	-		80.00 MHz (int. Å) 64.00 MHz (int. B)		1 🔽 167772	215 🚽 🔽 1		
	Locked if either physical key switched		50.00 MHz (int. C)		2 🔽 🛛 7	18 🗐 🗌 2		
	or soft lock active!				3	0 🚍 📃 3		
					4 🔽 205	519 🚽 🗖 4		
	Laser Soft Lock				5 🗆	0 🚍 🗖 5		
					6 🗆	0 🔤 🗖 6		
	Laser Locked				7 🔽 7920916 🚽 🗖 7		[pulses] [msec]	
					8 🔽 88350	62 🚍 🗖 8	Sequencer period:	33554430 := 419.4
1	Slot 200			Slot 500	Slot 600	Slot 700	Slot 800	Slot 900
Y	SLM 828	SLM 828	SLM 828	SLM 828	SLM 828	SLM 828	SLM 828	SLM 828
Int	ensitv	Intensity	Intensity	Intensity	Intensity	Intensity	Intensity	Intensity
	ero 100 ÷ %	Zero 100 ÷ %	Zero 100 + %	Zero 80 ÷ %	Zero 60 ÷ %	Zero 100 ÷ %		Zero 80 ÷ %
	igger	Trigger	Trigger	Trigger	Trigger	Trigger	Trigger	Trigger
ri	sing (ext.) 💌	rising (ext.) 💌	rising (ext.) 💌	rising (ext.) 💌	rising (ext.) 💌	rising (ext.) 💌	rising (ext.) 💌	rising (ext.) 💌
C	ont. Wave/	80 MHz (int.) 40 MHz (int.)	Cont. Wave/	Cont. Wave/ Pulse Disable	Cont. Wave/ 🗆 Pulse Disable	Cont. Wave/ Pulse Disable	Cont. Wave/ 🖂 Pulse Disable	Cont. Wave/
P	ulse Disable	20 MHz (int.) 10 MHz (int.)	Pulse Disable	Pulse Disable '	Pulse Disable '	Pulse Disable '	Pulse Disable '	Pulse Disable
	laser	5 MHz (int.) 2.5 MHz (int.)	laser	laser	laser	laser	laser	laser
		rising (ext.) falling (ext.)				[	1	
	Select Device	(onc)			Discard	Apply		<u>Q</u> uit

Screenshot from "Sepia II" control software. Please see text for associations 1 to 4 shown above.

# Laser driver module - SLM 828 1

A laser driver module is necessary for each laser head that needs to be connected and controlled at the same time by the PDL 828 "Sepia II". Besides the connector for the laser head, the module provides its own synchronization output, a trigger input and two gating inputs (fast and slow gate).

## **Triggering modes**

Each laser driver module can operate either independently from the other modules driven by its internal oscillator at six fixed, user-selectable frequencies (80, 40, 20, 10, 5 or 2.5 MHz) or synchronized to each other when using the external trigger signal from the oscillator module or any other source of a NIM compatible signal.


### **Emission intensity**

The intensity of the laser emission is controlled via an internal voltage in steps of 1 % of its full scale value. However, due to the non-linear behaviour of laser diodes, the voltage setting does not linearly correspond to the output power, i.e. a voltage of 50 % does not correspond to 50 % of the maximum output power.

A special feature of the PDL 828 "Sepia II" is to switch between cw and pulsed mode with the LDH-D-C Series of laser heads.

### **Gating functions**

For special applications like scanning devices, the laser driver module has two gating functions which allow to suppress the laser emission by an external signal: A slow gate, that reduces setting times of the laser heads to a minimum at slow on/off periods (seconds) and a fast gate, that can perform transitions within nanoseconds, i.e. in between two pulses. The fast gate also provides high pulse stability when the on/off signal is periodic and fast (milliseconds).




## **Oscillator module - SOM 828**

The oscillator module has eight (individual) trigger outputs (channels) that can each be addressed individually. For example, channels can be combined to be activated at the same time or each channel can be individually activated in a sequence. To enable burst operation, multiple pulses can be output from one channel (or from combined channels) before the next channel becomes active. Additionally, the oscillator has its own synchronization output, an external trigger input as well as an auxiliary input and output connector.

## Working principle

Generally speaking, the oscillator module works in a rotary fashion, which means that the programmed sequence of channel 1 through channel 8 must be completed before channel 1 becomes active again as illustrated in the figure below.

All channels can therefore be activated individually in a sequence. Adjacent channels can also be combined for simultaneous operation. The channels can, however, not be combined in any arbitrary fashion as the rotary working principle must be preserved.



The auxiliary input can be used to inhibit the oscillator to start the output period, i.e. to "start" the output on channel 1. If the signal is provided during the rotation

through the channels, one full period will be finished nonetheless.

The auxiliary output is used to signal the start of a period. It is active when a period is finished (i.e. if channel 8 has been active) and a new one is about to start. This special time is defined as the moment, when the first pulse is output from channel 1.

The oscillator works in a rotary fashion, i.e. the channels (C) are activated in sequence: 1, 2, 3, 4, 5, 6, 7, 8, 1, 2, 3, ...

## Repetition rate 2

The oscillator module of the "Sepia II" provides a wide range of userselectable repetition rates to generate pulses at the eight output channels of the module. In principle, all repetition rates are derived from an internal crystal oscillator along with a frequency divider. The oscillator module has three internal user-selectable crystal oscillators with base frequencies of 80, 64 and 50 MHz. Each base frequency can be further reduced by division though any integer value between 1 and 255. The highest repetition rate is therefore 80 MHz and the lowest repetition rate is 50 MHz/255~196 kHz. Instead of using one of the three internal oscillators, it is also possible to provide an external trigger signal. The frequency divider is also active on external trigger signals thereby allowing virtually all repetition rates between single shot and 80 MHz.

### Defining bursts and combining channels (3)

The oscillator module allows to output any number of pulses between 1 and 16772215 (16.7 million) to one channel before the next channel becomes active ("bursts"). Adjacent output channels can also be combined to emit pulses at the same time.

The output of each channel can be enabled or disabled. This does, however, not mean that the channel is "eliminated" from the rotary working principle. Instead, the programmed number of pulses is still processed, but simply no signal is present at the output. This is useful to insert time gaps between two bursts.


#### The synchronization signal 4

Each pulse that is output at any of the eight channels can be accompanied by a synchronization signal at the sync out connector of the oscillator module. This synchronization signal can be used to control external hardware like e.g. provide the start pulse for a TCSPC system. It can be enabled and disabled individually for each channel. One special feature of the PDL 828 "Sepia II" is that the synchronization signal can be time shifted at lower repetition frequencies. If the divider ratio is > 1, the sync pulse can be moved within the main clock period (the divided signal) in steps of the crystal oscillator's period (pre-sync). If the pre-sync is set to a value equal or higher than the divider, the synchronization signal will be disabled.


An additional option to influence the synchronization signal is to mask a defined number of pulses. Masking in this context means that no synchronization signals are output. The number of omitted pulses can be set to any integer value between 0 and 255.

The sync mask can also be inverted, i.e. not the number of omitted pulses is specified, but the number of actual output pulses. Again, the number of inverted masked synchronization pulses can be set to any integer value between 0 and 255. Of course, a value of 0 leads to no synchronization output at all.


# Pulse pattern examples




The repetition rate is derived from an internal crystal oscillator by division through any integer value between 1 and 255. If, for example, the base frequency (F) of 80 MHz is selected, a divider setting (D) of 2 yields a repetition rate of 40 MHz and a divider setting of 5 a repetition rate of 16 MHz.



By using different burst lengths it is possible to generate numerous pulse patterns. In this example, channel 1 and channel 2 are combined (burst length - B - of channel 2 is set to zero). Therefore, channel 1 and channel 2 emit three pulses at the same time, followed by six pulses out of channel 3.



The synchronization signal (S) can be output in advance to the output of each channel. The corresponding time difference (presync P) can be changed in the stepsize of the crystal oscillator's period. If the pre-sync is set to a value equal or greater than the divider setting, the synchronization signal will be disabled.



The synchronization signal (S) can be masked (M) or inverted masked (iM), i.e. omitted resp. allowed for a defined number of pulses. As a result, pulses at the beginning resp. at the end of the burst are not accompanied by a synchronization signal.

# Specifications

Mainframe         Large, L       1 slot for oscillator module, 8 slots for laser driver modules         Small, S       1 slot for oscillator module, 2 slots for laser driver modules         Power supply       100 to 250 VAC, 50/60 Hz, max. 250 Watts (large) or 130 Watts (small)         Dimensions       1arge, L: 464 × 310 × 140 mm (w × d × h)         small, S: 250 × 310 × 140 mm (w × d × h)
Oscillator module         Outputs       8 trigger (NIM), 1 synchronization (NIM), 1 auxiliary         Inputs       1 external trigger, 1 auxiliary (TTL)         Operation mode       rotary, programmed sequence of one channel must be completed before next         channel is activated, multiple channels can be combined         Oscillator type       crystal locked         Base frequencies       80, 64, 50 MHz (selectable)         Repetition frequency       user-selectable, derived from the selected base frequency or an external trigger source by division through any integer factor between 1 and 256
Synchronization output         Timing       synchronous to repetition frequency, timing position stepwise adjustable within the limits of the repetition frequency, stepsize equals base oscillator period         Masking       synchronization pulses can be inhibited (masked), mask size selectable in integer steps from 0 to 255, stepsize equals repetition period         Amplitude       < -800 mV into 50 Ohms (NIM)
Auxiliary output         Timing       at start of complete trigger sequence         Amplitude       +500 mV into 50 Ohms
External trigger input Amplitude5 to +5 V (maximum limits) Trigger level1 to +1 V Frequency rangeup to 40 MHz
<i>Bursts</i> Burst length
Laser driver module         Inputs
Computer Operating system

Further available are Fluorescence Lifetime Spectrometer; Time-resolved Fluorescence Microscopes; Upgrade kit for Laser Scanning Microscopes; Picosecond / Nanosecond Pulsed, Modulated and Fast Swichted Diode Lasers; PC Modules for TCSPC. Please call for detailed information and data sheets. OEM Modules of all products are available upon request. Please check our website for updated information.



For all available types and wavelengths please go to: http://www.picoquant.com/products/category/picosecond-pulsed-sources



INVISIBLE OR VISIBLE LASER RADIATION VOID DIRECT EXPOSURE TO BEAM CLASS 3B LASER PRODUCT IEC / EN 60825-1

All Information given here is reliable to our best knowledge. However, no responsibility is assumed for possible inaccuracies or omissions. Specifications and external appearances are subject to change without notice. Trademarks or corporate names are used for explanation and identification, to the owner's benefit and without intent to infringe.



PicoQuant GmbH Rudower Chaussee 29 (IGZ) 12489 Berlin Germany Phone +49-(0)30-6392-6929 Telefax +49-(0)30-6392-6561 Email info@picoquant.com WWW http://www.picoquant.com